

@ML521_SPE

ANALYTE	METHOD	CONTAINER & PRESERVATIVES	HOLDING TIME	MRL	MDL	UNITS	LCS RANGE	MS RANGE	CAS #
N-Nitroso dimethylamine (NDMA)	EPA 521	1L amber glass 80-100mg Sodium Thiosulfate	28 DAY	2	0.962	ng/L	70 - 130	70 - 130	62-75-9
N-Nitrosodibutylamine (NDBA)	EPA 521	1L amber glass 80-100mg Sodium Thiosulfate	28 DAY	2	1.569	ng/L	70 - 130	70 - 130	
N-Nitrosodiethylamine (NDEA)	EPA 521	1L amber glass 80-100mg Sodium Thiosulfate	28 DAY	2	1.493	ng/L	70 - 130	70 - 130	55-18-5
N-Nitrosodi-n-propylamin(NDPA)	EPA 521	1L amber glass 80-100mg Sodium Thiosulfate	28 DAY	2	0.843	ng/L	70 - 130	70 - 130	
N-Nitrosomethylethylamin(NMEA)	EPA 521	1L amber glass 80-100mg Sodium Thiosulfate	28 DAY	2	0.843	ng/L	70 - 130	70 - 130	
N-Nitrosopyrrolidine (NPYR)	EPA 521	1L amber glass 80-100mg Sodium Thiosulfate	28 DAY	2	0.936	ng/L	70 - 130	70 - 130	930-55-2